



All dimensions are in mm; tolerances according to ISO 2768 m-H

**Interface**

According to  
Mechanically compatible with

IEC 61169-23  
RPC-2.92 and SMA

**Contents and Documentation**

This kit is delivered with

- **Standard Definitions Card**  
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers
- **Test Results Documentation**
- **Hard Shell Case**

**Material and plating**

**Connector parts**

Center conductor  
Outer conductor  
Body  
Dielectric  
Substrate

**Material**

Beryllium copper  
Stainless steel  
Aluminum  
PS  
Al<sub>2</sub>O<sub>3</sub>

**Plating**

Gold, min. 1.27 µm, over nickel  
Passivated  
black anodized

**Electrical data**

Frequency range DC to 26.5 GHz

**Open**

Error from nominal phase<sup>1</sup>  
 $\leq 1.0^\circ$ , DC to 4 GHz  
 $\leq 2.0^\circ$ , 4 GHz to 8 GHz  
 $\leq 3.0^\circ$ , 8 GHz to 26.5 GHz

**Short**

Error from nominal phase<sup>2</sup>  
 $\leq 1.0^\circ$ , DC to 4 GHz  
 $\leq 2.0^\circ$ , 4 GHz to 8 GHz  
 $\leq 3.0^\circ$ , 8 GHz to 26.5 GHz

**Load**

Return loss  
 $\geq 40.0$  dB, DC to 4 GHz  
 $\geq 35.0$  dB, 4 GHz to 8 GHz  
 $\geq 30.0$  dB, 8 GHz to 26.5 GHz

DC Resistance 50  $\Omega \pm 0.5 \Omega$   
 Power handling  $\leq 0.5$  W

<sup>1</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Fringing Capacitances

<sup>2</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Short Inductance

**Mechanical data**

Mating cycles  $\geq 500$   
 Maximum torque 1.70 Nm  
 Recommended torque 0.90 Nm  
 Gauge 0.00 mm to 0.08 mm

**General standard definitions**

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

**Open**

Offset  $Z_o$  / Impedance /  $Z_o$  50  $\Omega$   
 Offset Delay 33.356 ps  
 Length (electrical) / Offset Length 10.00 mm  
 Offset Loss 2.20 G $\Omega$ /s  
 Loss 0.0127 dB/  $\sqrt{\text{GHz}}$   
 Fringing Capacitances  
 $C_0 = -17.500 \times 10^{-15}$  F / -17.500 fF  
 $C_1 = -2000.0 \times 10^{-27}$  F/Hz / -2.0000 fF /GHz  
 $C_2 = 140.00 \times 10^{-36}$  F/Hz / 0.1400 fF /GHz<sup>2</sup>  
 $C_3 = -2.7000 \times 10^{-45}$  F/Hz / -0.0027 fF /GHz<sup>3</sup>

**Short**

Offset $Z_0$ / Impedance / $Z_0$	50 $\Omega$	
Offset Delay	33.356 ps	
Length (electrical) / Offset Length	10.00 mm	
Offset Loss	2.36 G $\Omega$ /s	
Loss	0.0137 dB/ $\sqrt{\text{GHz}}$	
Short Inductance	$L_0 = -44.000 \times 10^{-12} \text{ H}$	/ -44.000 pH
	$L_1 = 3700.0 \times 10^{-24} \text{ H/Hz}$	/ 3.7000 pH /GHz
	$L_2 = -250.00 \times 10^{-33} \text{ H/Hz}^2$	/ -0.2500 pH /GHz <sup>2</sup>
	$L_3 = 5.0000 \times 10^{-42} \text{ H/Hz}^3$	/ 0.0050 pH /GHz <sup>3</sup>

**Load**

Offset $Z_0$ / Impedance / $Z_0$	50 $\Omega$
Offset Delay	0.0000 ps
Length (electrical) / Offset Length	0.000 mm
Offset Loss	0.00 G $\Omega$ /s
Loss	0.0000 dB/ $\sqrt{\text{GHz}}$

**Environmental data**

Operating temperature range <sup>3</sup>	+20 °C to +26 °C
Rated temperature range of use <sup>4</sup>	0 °C to +50 °C
Storage temperature range	-40 °C to +85 °C
RoHS	compliant

<sup>3</sup> Temperature range over which these specifications are valid.

<sup>4</sup> This range is underneath and above the operating temperature range, within the calibration kit is fully functional and could be used without damage

**Declaration of documentation**

Standard delivery for this kit includes Test Results. The documentation issued reports which quantities were tested individually, traceable to national / international standards. Model based standard definitions of the calibration standards are reported in Agilent / Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

**Inspection interval**

Recommendation 12 months

**Packing**

Standard 1 pce in bag  
Weight 24.7 g/pce

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Marcel Panicke	07.01.16	Markus Müller	02.05.18	d00	18-0787	Marion Striegler	02.05.18
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de						Tel. : +49 8684 18-0 Email : info@rosenberger.de	
							Page 3 / 3